



Figure S3: Explanation for premature valving. If a liquid plug is unintentionally present in the venting channel, an enclosed air volume (light red) is expanded during temperature increase of the hot start for PCR. As a result, the liquid (dark blue) is displaced through the siphon valve (light blue) during pre-amplification.